

Errata for “USB Device Class Definition for Personal Healthcare Devices, Release 1.0, November 8, 2007” as of February 15, 2008.

Check the <http://www.usb.org> website for the latest errata.

Chapter 7

Section 7.1.2

Background: The Get Data Status request has a *wLength* of two. However, it is incorrectly displayed as zero below Table 14.

Change: Change “If *wLength* is non-zero, then the device behavior is not specified.” to “If *wLength* is a value other than two (2), then the device behavior is not specified.”

Chapter 8

Table 17

Background: FEATURE_PHDC_METADATA is incorrectly listed as FEATURE_PHDC_QOS.

Change: Change FEATURE_PHDC_QOS to FEATURE_PHDC_METADATA.